8650A Series Universal Power Meters





8650A Series Universal Power Meters

The Capabilities to Test Today's Sophisticated Communications Systems

The Giga-tronics 8650A Series Universal Power Meters have the extensive measurement capabilities and unique features required to test today's sophisticated communications systems faster and more accurately.

TDMA

The 8650A can automatically measure the average power of pulse modulated signals or pulse signals that are ampli-



tude modulated during the pulse 'on' period — such as TDMA signals.

Using the exclusive Burst Average Power mode (BAP), the average power reading in the pulse burst is automatically measured between the 3 dB points. Therefore, the duty cycle can change in time without affecting the accuracy of the meter reading. This method eliminates the need to manually set time gating, which can add errors if the gate is not set accurately.

GSM

The Time Gating feature of the 8650A lets you program





The Giga-tronics 8650A Series combines the speed, range and capabilities needed to test today's sophisticated communications systems.

Giga-tronics 8650A Features and Specifications

a measurement start time and duration to measure the average power during a specific time period of a GSM burst signal. The graphic display provides visual feedback if you prefer to set the gate manually. And, of course, there is the ability to use the TTL signal for automatically setting the time gate control.

CDMA

The 8650A has the wide, 80 dB single sensor dynamic range required for CDMA signal open-loop tests, the speed you need to quickly measure power during closed-loop tests, and the 10 MHz bandwidth needed to test third-generation CDMA signals.



INSTANTANEOUS PEAK POWER

You can also measure the instantaneous peak power level of a pulse modulated signal with the 8650A.



A built-in delay line lets you trigger a few nanoseconds ahead of the pulse for rising edge measurements. While a built-in time base gives you sample delay control up to 100 ms after the trigger point with 0.5 ns resolution. And you can view the profile and see the exact measurement point on the pulse.

MAXIMUM PEAK POWER

The peak hold feature of the

GPIB CW Measurement Speed (rdgs/s)	
Normal Mode	>300
Swift Mode	>1,750
Fast Buffered Mode	>26,000
GPIB Modulated Measurement Speed (rdgs/s)	
Normal Mode	>150
Swift Mode	>800
Fast Modulated Mode	>800
Asynchronous Sample Rate	2.5—5 MHz
Maximum Diode Sensor Video Bandwidth	20 MHz
Maximum Instrument Video Bandwidth	10 MHz
Maximum Single Sensor CW Dynamic Range	90 dB
Maximum Single Sensor Modulation Dynamic Range	
TDMA/GSM	60—80 dB
CDMA (IS-95)	80 dB
Wideband CDMA (10 MHz bandwidth)	80 dB
Maximum Peak Power Sensor Rise Time	100 ns
Automatic Time Gate Setting	Yes
Direct Crest Factor Measurement	Yes
Statistical Power Measurement Analysis	Yes

8650A lets you display the highest instantaneous power measured from the time the feature is enabled until it is reset.

The display value tracks the measured value only



when it is rising to a new maximum; when the measured value falls, the display value holds at the maximum.



CREST FACTOR

The crest factor capability of the 8650A displays the ratio of the maximum peak power (peak hold) measurement to the average power measurement (in dB) from the time the feature is enabled until it is reset.

The crest factor capability operates in the same manner as the peak hold capability: the display value holds at the maximum until it is reset.

INCREDIBLE SPEED AND STATISTICAL ANALYSIS

No other meter delivers the measurement speed available from the 8650A.

Achieve over 1,750 readings per second over GPIB. HOLD exclusive fast buffered mode to further

Or use our

reduce processor overhead and capture over 26,000 readings per second.

Incredible speed for CW and modulated measurements results from an asynchronous

sampling rate of 2.5 to 5 MHz, that minimizes the aliasing effects of signals to produce faster average power measurements.

And the 8650A features a wide variety of statistical

power measurement analysis, to evaluate communications system efficiency.

BURST START AND END EXCLUDE

The exclusive burst start and end exclude capabilities of the 8650A allow you to exclude the beginning or end of a burst when measuring the average burst power.



Masking the beginning or the end of a burst signal, in order to exclude overshoot or other distortions, can be desirable or even required for certain types of power measurements.



Unrivaled Accuracy and Built-In Calibration

Giga-tronics uses diode sensors exclusively to provide speed, range, capability and accuracy unavailable from any other power meter.

ACCURACY OVER A 90 dB RANGE

Giga-tronics has solved the problem that limited the use of diode sensors to below -20 dBm — the 'square law' region — by utilizing a patented built-in power sweep calibration system.

The power sweep calibrator uses a 50 MHz amplitude controlled oscillator to step from –30 to +20 dBm in 1 dB increments. Each step is set using an internal thermistor — the standard for accuracy and traceability.

Giga-tronics gives you thermistor accuracy plus diode speed for measuring signals over a full 90 dB power range.

BUILT-IN FREQUENCY RESPONSE CALIBRATION

Configuring the meter for measurements is easy with calibration factors programmed into the sensor.

When the measurement frequency is entered, the meter automatically applies the correct calibration factor from the sensor EEPROM. And the meter automatically

reads a new set of cal factors when a sensor is changed.

This avoids the chance of measurement error from using invalid calibration factors when you change sensors, or from forgetting to

Accuracy Audit

The Accuracy Audit table lists the significant uncertainties of an absolute power measurement. The accuracy of the 8650A combined with the 80301A sensor is compared to a typical thermocouple sensor/meter combination at +20 dBm, 0 dBm, and -30 dBm (the dynamic limit of the thermocouple sensor). The uncertainty comparison at -30 dBm illustrates the accuracy advantage of a wide dynamic sensor, even when the full 90 dB dynamic range is not utilized.

Frequency = 1 GHz; Source Match = 1.5.1with 80301AThermocouple Meter/SensorInstrumentation Uncertainty $\pm 5.2\%$ $\pm 2.5\% - 4.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibration Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.0000005\%$ $\pm 0.000005\%$ Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ $\%$ Total Uncertainty $\pm 9.97\%$ $\pm 7.53 - 9.53\%$ dB Total Uncertainty ± 0.41 dB $\pm 0.316 - 0.4$ dBO dBm Frequency = 1 GHz; Sourcewith with Match = 1.5:1Thermocouple With Meter/SensorInstrumentation Uncertainty $\pm 0.5\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GH2) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 0.23\%$ Calibrator Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator Source) $\pm 2.25\%$ $\pm 2.0\%$ $\%$ Total Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ $\%$ Total Uncertainty $\pm 0.22\%$ $\pm 0.23\%$ Match = 1.5:18650ATypical Thermocouple WithMismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ $\%$ Total Uncertainty ± 0.22 ± 0.23 $\%$ Total Uncertainty $\pm 0.22\%$ $\pm 0.5\%$	Frequency = 1 GHz; Source Match = 1.5:1with 80301AThermocouple Meter/SensorInstrumentation Uncertainty $\pm 5.2\%$ $\pm 2.5\%$ $\pm 2.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.0000005\%$ $\pm 0.0000005\%$ Noise $\pm 0.0000005\%$ $\pm 0.0001\%$ Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ $\%$ Total Uncertainty $\pm 9.97\%$ ∓ 7.53 -9.53%dB Total Uncertainty ± 0.41 dB ± 0.316 - 0.4 dBO dBm Frequency = 1 GHz; Source Match = 1.5:18650A with 80301ATypical Meter/SensorInstrumentation Uncertainty $\pm 0.5\%$ $\pm 0.23\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Inserver Mismatch $\pm 0.28\%$ $\pm 0.005\%$ Noise $\pm 0.000005\%$ $\pm 0.00\%$ Galibrator Uncertainty $\pm 0.225\%$ $\pm 0.23\%$ Calibrator Uncertainty $\pm 0.225\%$ $\pm 0.23\%$	+20 dBm	8650A	Typical
Match = 1.5:180301AMeter/SensorInstrumentation Uncertainty $\pm 5.2\%$ $\pm 2.5\% - 4.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibration Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.0000005\%$ $\pm 0.000005\%$ Noise $\pm 0.0000005\%$ $\pm 0.0001\%$ Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ $\%$ Total Uncertainty $\pm 9.97\%$ $\pm 7.53 - 9.53\%$ dB Total Uncertainty $\pm 9.97\%$ $\pm 7.53 - 9.53\%$ dB Total Uncertainty ± 0.41 dB $\pm 0.316 - 0.4$ dBO dBm8650ATypicalFrequency = 1 GHz; SourcewithThermocoupleMatch = 1.5:180301AMeter/SensorInstrumentation Uncertainty $\pm 0.5\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHZ) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator Source) $\pm 2.25\%$ $\pm 2.0\%$ $\%$ Total Uncertainty $\pm 0.22\%$ $\pm 0.23\%$ Noise $\pm 0.000005\%$ $\pm 0.005\%$ Noise $\pm 0.000005\%$ $\pm 0.02\%$ $\%$ Total Uncertainty $\pm 5.27\%$ $\pm 5.54\%$ $Match = 1.5:1$ 80301AMeter/SensorInstrumentation Uncertainty $\pm 0.925\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0.5\%$ Sensor Pow	Match = 1.5:1 80301A Meter/Sensor Instrumentation Uncertainty $\pm 5.2\%$ $\pm 2.5\% - 4.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.00000005\%$ $\pm 0.0000005\%$ Noise $\pm 0.00000005\%$ $\pm 0.0001\%$ Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ % Total Uncertainty $\pm 9.97\%$ $\pm 7.53 - 9.53\%$ dB Total Uncertainty ± 0.41 dB $\pm 0.316 - 0.4$ dB O dBm 8650A Typical Frequency = 1 GHz; Source with Thermocouple Match = 1.5:1 80301A Meter/Sensor Instrumentation Uncertainty $\pm 0.28\%$ $\pm 0.23\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.005\%$ Calibrator/Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$	Frequency = 1 GHz; Source	with	Thermocouple
Instrumentation Uncertainty $\pm 5.2\%$ $\pm 2.5\%$ $\pm 2.0\%$ $\pm 0.23\%$ $\equiv 0.23\%$ $\equiv 0.23\%$ $\equiv 0.23\%$ $\equiv 0.23\%$ $\equiv 0.000005\%$ $\pm 0.00005\%$ $\pm 0.00005\%$ $\pm 0.00005\%$ $\pm 0.000005\%$ $\pm 0.000005\%$ $\pm 0.000005\%$ $\pm 0.000005\%$ $\pm 0.00005\%$ $\pm 0.000005\%$ $\pm 0.01\%$ $\pm 0.5\%$ ± 0.51 ± 0.51 ± 0.51 ± 0.51 $\pm 0.52\%$ $\pm 2.0\%$ $\pm 0.55\%$ $\pm 0.0000005\%$ $\pm 0.05\%$ $\pm 0.000005\%$ $\pm 0.05\%$ $\pm 0.05\%$ $\pm 0.55\%$ $\pm 0.000005\%$ $\pm 0.05\%$ $\pm 0.000005\%$ $\pm 0.005\%$ $\pm 0.000005\%$ $\pm 0.005\%$ $\pm 0.000005\%$ $\pm 0.005\%$ $\pm 0.000005\%$ $\pm 0.000005\%$ $\pm 0.005\%$ $\pm 0.000005\%$ $\pm 0.000005\%$ $\pm 0.0005\%$ $\pm 0.000005\%$ $\pm 0.0005\%$ $\pm 0.000005\%$ $\pm 0.00005\%$ $\pm 0.005\%$ $\pm 0.000005\%$ $\pm 0.000005\%$ $\pm 0.0005\%$ $\pm 0.000005\%$ $\pm 0.005\%$ $\pm 0.000005\%$ $\pm 0.0000005\%$	$\begin{array}{rrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrr$	Match = 1.5:1	80301A	Meter/Sensor
Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibration Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.0000005\%$ $\pm 0.00005\%$ Noise $\pm 0.0000005\%$ $\pm 0.0001\%$ Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ $\%$ Total Uncertainty $\pm 9.97\%$ $\pm 7.53 - 9.53\%$ dB Total Uncertainty $\pm 9.97\%$ $\pm 7.53 - 9.53\%$ dB Total Uncertainty ± 0.41 dB $\pm 0.316 - 0.4$ dBO dBm8650ATypicalFrequency = 1 GHz; SourcewithThermocoupleMatch = 1.5:180501AMeter/SensorInstrumentation Uncertainty $\pm 0.5\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0.23\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator Incertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.000005\%$ $\pm 0.005\%$ Noise $\pm 0.00005\%$ $\pm 0.01\%$ Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ $\%$ Total Uncertainty $\pm 5.27\%$ $\pm 5.54\%$ Hereucey = 1 GHz; SourcewithThermocoupleMatch = 1.5:180301AMeter/SensorInstrumentation Uncertainty $\pm 0.925\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0\%$	Sensor Power Linearity (>8 GHz) \pm 0% \pm 0% Calibrator Uncertainty \pm 1.2% \pm 1.2% Calibrator/Sensor Mismatch \pm 0.28% \pm 0.23% Calibration Factor Uncertainty \pm 1.04% \pm 1.6% Zero Set \pm 0.00000005% \pm 0.0001% Noise \pm 0.00000005% \pm 0.0001% Mismatch (Sensor/Source) \pm 2.25% \pm 2.0% % Total Uncertainty \pm 9.97% \pm 7.53 -9.53% dB Total Uncertainty \pm 9.97% \pm 7.53 -9.53% dB Total Uncertainty \pm 0.41 dB \pm 0.316 - 0.4 dB 0 dBm 8650A Typical Frequency = 1 GHz; Source with Thermocouple Match = 1.5:1 80301A Meter/Sensor Instrumentation Uncertainty \pm 0.28% \pm 0.23% Calibrator/Sensor Mismatch \pm 0.28% \pm 0.23% Calibrator Uncertainty \pm 1.04% \pm 1.6% Zero Set \pm 0.000005% \pm 0.01% Mismatch (Sensor/Source) \pm 2.25% \pm 2.0% % Tota	Instrumentation Uncertainty	±5.2%	+ 2.5% -4.5%
Calibrator Uncertainty \pm 1.2% \pm 1.2%Calibrator/Sensor Mismatch \pm 0.28% \pm 0.23%Calibration Factor Uncertainty \pm 1.04% \pm 1.6%Zero Set \pm 0.0000005% \pm 0.000005%Noise \pm 0.0000005% \pm 0.0001%Mismatch (Sensor/Source) \pm 2.25% \pm 2.0%% Total Uncertainty \pm 9.97% \pm 7.53 -9.53%dB Total Uncertainty \pm 9.97% \pm 7.53 -9.53%dB Total Uncertainty \pm 0.41 dB \pm 0.316 - 0.4 dB0 dBm8650ATypicalFrequency = 1 GHz; SourcewithThermocoupleMatch = 1.5:180301AMeter/SensorInstrumentation Uncertainty \pm 0.5% \pm 0.5%Sensor Power Linearity (>8 GHz) \pm 0% \pm 0.23%Calibrator/Sensor Mismatch \pm 0.28% \pm 0.23%Calibrator/Sensor Mismatch \pm 0.28% \pm 0.005%Noise \pm 0.000005% \pm 0.01%Mismatch (Sensor/Source) \pm 2.25% \pm 2.0%% Total Uncertainty \pm 5.27% \pm 5.54%dB Total Uncertainty \pm 0.22 dB \pm 0.23 dB-30 dBm8650ATypicalFrequency = 1 GHz; SourcewithThermocoupleMatch = 1.5:18650ATypicalCalibrator Uncertainty \pm 0.925% \pm 0.5%Sensor Power Linearity (>8 GHz) \pm 0% \pm 0.0%Calibrator Uncertainty \pm 0.28% \pm 0.23%Calibrator Uncertainty \pm 0.28% \pm 0.23%Calibrator Uncertainty	$\begin{array}{llllllllllllllllllllllllllllllllllll$	Sensor Power Linearity (>8 GHz)	± 0%	± 0%
Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibration Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.0000005\%$ $\pm 0.00005\%$ Noise $\pm 0.0000005\%$ $\pm 0.0001\%$ Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ % Total Uncertainty $\pm 9.97\%$ ± 7.53 -9.53%dB Total Uncertainty $\pm 9.97\%$ ± 7.53 -9.53%dB Total Uncertainty ± 0.41 dB ± 0.316 - 0.4 dB0 dBm8650ATypicalFrequency = 1 GHz; SourcewithThermocoupleMatch = 1.5:180301AMeter/SensorInstrumentation Uncertainty $\pm 0.5\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator Incertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.00005\%$ $\pm 0.005\%$ Noise $\pm 0.00005\%$ $\pm 0.01\%$ Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ % Total Uncertainty $\pm 5.27\%$ $\pm 5.54\%$ Match = 1.5:180301AMeter/SensorInstrumentation Uncertainty $\pm 0.925\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 0.23\%$ Ca	$\begin{array}{llllllllllllllllllllllllllllllllllll$	Calibrator Uncertainty	± 1.2%	± 1.2%
Calibration Factor Uncertainty \pm 1.04% \pm 1.6% Zero Set \pm 0.0000005% \pm 0.000005% Noise \pm 0.00000005% \pm 0.0001% Mismatch (Sensor/Source) \pm 2.25% \pm 2.0% % Total Uncertainty \pm 9.97% \pm 7.53 –9.53% dB Total Uncertainty \pm 0.41 dB \pm 0.316 – 0.4 dB O dBm 8650A Typical Frequency = 1 GHz; Source with Thermocouple Match = 1.5:1 80301A Meter/Sensor Instrumentation Uncertainty \pm 0.5% \pm 0.5% Sensor Power Linearity (>8 GHz) \pm 0% \pm 0% Calibrator /Sensor Mismatch \pm 0.28% \pm 0.23% Calibrator Sensor Mismatch \pm 0.28% \pm 0.005% Noise \pm 0.000005% \pm 0.01% Mismatch (Sensor/Source) \pm 2.25% \pm 2.0% % Total Uncertainty \pm 5.27% \pm 5.54% dB Total Uncertainty \pm 0.22 dB \pm 0.23 dB -30 dBm 8650A Typical Frequency = 1 GHz; Source with Thermocouple Match = 1.5:1	$\begin{array}{rrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrr$	Calibrator/Sensor Mismatch	± 0.28%	± 0.23%
Zero Set $\pm 0.0000005\%$ $\pm 0.000005\%$ Noise $\pm 0.0000005\%$ $\pm 0.0001\%$ Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ % Total Uncertainty $\pm 9.97\%$ $\pm 7.53 - 9.53\%$ dB Total Uncertainty $\pm 9.97\%$ $\pm 7.53 - 9.53\%$ dB Total Uncertainty ± 0.41 dB $\pm 0.316 - 0.4$ dB O dBm 8650A Typical Frequency = 1 GHz; Source with Thermocouple Match = 1.5:1 80301A Meter/Sensor Instrumentation Uncertainty $\pm 0.5\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0.\%$ $\pm 0.\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.00005\%$ $\pm 0.005\%$ Noise $\pm 0.00005\%$ $\pm 0.01\%$ Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ % Total Uncertainty ± 0.22 dB ± 0.23 dB -30 dBm 8650A Typical Frequency = 1 GHz; Source wit	$\begin{array}{rrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrr$	Calibration Factor Uncertainty	± 1.04%	± 1.6%
Noise \pm 0.0000005% \pm 0.0001%Mismatch (Sensor/Source) \pm 2.25% \pm 2.0%% Total Uncertainty \pm 9.97% \pm 7.53 –9.53%dB Total Uncertainty \pm 0.41 dB \pm 0.316 – 0.4 dB0 dBm8650ATypicalFrequency = 1 GHz; SourcewithMatch = 1.5:1Meter/SensorInstrumentation Uncertainty \pm 0.5% \pm 0.5%Sensor Power Linearity (>8 GHz) \pm 0% \pm 0.23%Calibrator Uncertainty \pm 1.2% \pm 1.2%Calibrator/Sensor Mismatch \pm 0.28% \pm 0.23%Calibration Factor Uncertainty \pm 1.4% \pm 1.6%Zero Set \pm 0.000005% \pm 0.005%Noise \pm 0.000005% \pm 0.01%Mismatch (Sensor/Source) \pm 2.25% \pm 2.0%% Total Uncertainty \pm 5.27% \pm 5.54%dB Total Uncertainty \pm 0.22 dB \pm 0.23 dB-30 dBm8650ATypicalFrequency = 1 GHz; Sourcewith with Match = 1.5:1Instrumentation Uncertainty \pm 0.925% \pm 0.5%Sensor Power Linearity (>8 GHz) \pm 0% \pm 0.5%Sensor Power Linearity (>8 GHz) \pm 0% \pm 0.23%Calibrator/Sensor Mismatch \pm 0.28% \pm 0.23%Calibrator Uncertainty \pm 1.2% \pm 1.2%Calibrator Uncertainty \pm 1.2% \pm 1.2%Calibrator Uncertainty \pm 0.28% \pm 0.23%Calibrator Uncertainty \pm 0.28% \pm 0.23%Calibrator Uncertainty \pm 1.2%	$\begin{array}{rrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrr$	Zero Set	$\pm 0.0000005\%$	± 0.00005%
Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ % Total Uncertainty $\pm 9.97\%$ ± 7.53 dB Total Uncertainty ± 0.41 dB ± 0.316 0 dBm8650ATypicalFrequency = 1 GHz; SourcewithThermocoupleMatch = 1.5:180301AMeter/SensorInstrumentation Uncertainty $\pm 0.5\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GH2) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator Sensor Mismatch $\pm 0.28\%$ $\pm 0.005\%$ Noise $\pm 0.000005\%$ $\pm 0.005\%$ Noise $\pm 0.000005\%$ $\pm 0.01\%$ Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ $\%$ Total Uncertainty $\pm 5.27\%$ $\pm 5.54\%$ dB Total Uncertainty ± 0.22 dB ± 0.23 dB-30 dBm8650ATypicalFrequency = 1 GHz; SourcewithThermocoupleMatch = 1.5:180301AMeter/SensorInstrumentation Uncertainty $\pm 0.925\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GH2) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 0.23\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.005\%$ $\pm $	$\begin{array}{rrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrr$	Noise	$\pm 0.0000005\%$	± 0.0001%
% Total Uncertainty dB Total Uncertainty $\pm 9.97\%$ $\pm 0.41 dB$ $\pm 7.53 - 9.53\%$ $\pm 0.316 - 0.4 dB$ 0 dBm Frequency = 1 GHz; Source Match = 1.5:18650A with 80301ATypical Thermocouple Meter/SensorInstrumentation Uncertainty $\pm 0.5\%$ $\pm 0.5\%$ $\pm 0.5\%$ $\pm 0.5\%$ $\pm 0.5\%$ Calibrator Uncertainty (Calibrator Uncertainty) $\pm 1.2\%$ $\pm 1.2\%$ $\pm 1.2\%$ $\pm 0.23\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ $\pm 0.23\%$ $\pm 0.00005\%$ $\pm 0.005\%$ Noise $\pm 0.000005\%$ $\pm 0.00005\%$ $\pm 0.00005\%$ $\pm 0.005\%$ Noise $\pm 0.000005\%$ $\pm 0.00005\%$ $\pm 0.01\%$ Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ % Total Uncertainty $\pm 0.22 dB$ $\pm 0.23 dB$ -30 dBm Frequency = 1 GHz; Source Match = 1.5:18650A with Meter/SensorInstrumentation Uncertainty $\pm 1.2\%$ $\pm 0.925\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0.925\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 1.2\%$ $\pm 0.23\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 0.23\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 0.23\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.005\%$ $\pm 0.005\%$ Visio $\pm 0.005\%$ $\pm 0.005\%$ Calibrator Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.005\%$ $\pm 0.005\%$	$\begin{tabular}{ c c c c c } \hline \end{tabular} & $\pm 9.97\%$ & ± 7.53 - 9.53% \\ \hline \end{tabular} & ± 0.41 dB$ & ± 0.316 - 0.4 dB$ \\ \hline \end{tabular} & ± 0.41 dB$ & ± 0.316 - 0.4 dB$ \\ \hline \end{tabular} & ± 0.41 dB$ & ± 0.316 - 0.4 dB$ \\ \hline \end{tabular} & ± 0.41 dB$ & ± 0.316 - 0.4 dB$ \\ \hline \end{tabular} & ± 0.41 dB$ & ± 0.316 - 0.4 dB$ \\ \hline \end{tabular} & ± 0.41 dB$ & ± 0.316 - 0.4 dB$ \\ \hline \end{tabular} & ± 0.41 dB$ & ± 0.316 - 0.4 dB$ \\ \hline \end{tabular} & ± 0.41 dB$ & ± 0.316 - 0.4 dB$ \\ \hline \end{tabular} & ± 0.41 dB$ & ± 0.316 - 0.4 dB$ \\ \hline \end{tabular} & $\pm 0.5\%$ & $\pm 0.5\%$ \\ \hline \end{tabular} & $\pm 0.5\%$ & $\pm 0.5\%$ \\ \hline \end{tabular} & $\pm 0.5\%$ & $\pm 0.23\%$ & $\pm 0.23\%$ \\ \hline \end{tabular} & $\pm 0.000005\%$ \pm 0.005\%$ & $\pm 0.005\%$ \\ \hline \end{tabular} & $\pm 0.28\%$ \pm 0.23\%$ & $\pm 0.23\%$ \\ \hline \end{tabular} & $\pm 0.000005\%$ \pm 0.005\%$ \\ \hline \end{tabular} & $\pm 0.000005\%$ \pm 0.01\%$ \\ \hline \end{tabular} & $\pm 0.000005\%$ \pm 0.01\%$ \\ \hline \end{tabular} & $\pm 0.000005\%$ \pm 0.01\%$ \\ \hline \end{tabular} & ± 0.22 dB$ & ± 0.23 dB$ \\ \hline \end{tabular} & ± 0.23 dB$ \\ \hline \end{tabular} & ± 0.23 dB$ \\ \hline \end{tabular} & $\pm 0.925\%$ \pm 0.5\%$ \\ \hline \end{tabular} & $\pm 0.925\%$ \pm 0.5\%$ \\ \hline \end{tabular} & $\pm 0.925\%$ \pm 0.5\%$ \\ \hline \end{tabular} & $\pm 0.23\%$ dB$ \\ \hline \end{tabular} & $\pm 0.005\%$ \pm 10\%$ \\ \hline \end{tabular} & $\pm 0.25\%$ dB$ \\ \hline \end{tabular} & $\pm 0.005\%$ \pm 10\%$ \\ \hline \end{tabular} & $\pm 0.005\%$ \pm 10\%$ \\ \hline \end{tabular} & $\pm 0.25\%$ dB$ \\ \hline \end{tabular} & $\pm 0.25\%$ dB$ \\ \hline \end{tabular} & $\pm 0.25\%$ dB$	Mismatch (Sensor/Source)	± 2.25%	± 2.0%
dB Total Uncertainty \pm 0.41 dB+ 0.316 - 0.4 dB0 dBm Frequency = 1 GHz; Source8650A with B0301ATypical Thermocouple Meter/SensorInstrumentation Uncertainty \pm 0.5% \pm 0.5%Sensor Power Linearity (>8 GHz) \pm 0% \pm 0%Calibrator Uncertainty \pm 1.2% \pm 1.2%Calibrator Uncertainty \pm 1.2% \pm 1.2%Calibrator Uncertainty \pm 1.2% \pm 0.23%Calibrator Sensor Mismatch \pm 0.28% \pm 0.23%Calibrator Factor Uncertainty \pm 1.04% \pm 1.6%Zero Set \pm 0.000005% \pm 0.005%Noise \pm 0.000005% \pm 0.01%Mismatch (Sensor/Source) \pm 2.25% \pm 2.0%% Total Uncertainty \pm 5.27% \pm 5.54%dB Total Uncertainty \pm 0.22 dB \pm 0.23 dB-30 dBm8650A Mother/SensorTypical Meter/SensorInstrumentation Uncertainty \pm 0.925% \pm 0.5%Sensor Power Linearity (>8 GHz) \pm 0% \pm 0.0%Calibrator Uncertainty \pm 1.2% \pm 1.2%Calibrator Uncertainty \pm 1.2% \pm 0.23%Calibrator Uncertainty \pm 1.2% \pm 1.2%Calibrator Sensor Mismatch \pm 0.28% \pm 0.23%Calibrator Factor Uncertainty \pm 1.04% \pm 1.6%Zero Set \pm 0.005% \pm 5%Noise \pm 0.005% \pm 5%Noise \pm 0.005% \pm 5%Noise \pm 0.005% \pm 0%	$\begin{tabular}{lllllllllllllllllllllllllllllllllll$	% Total Uncertainty	±9.97%	+ 7.53 –9.53%
0 dBm8650A with 80301ATypical Thermocouple Meter/SensorInstrumentation Uncertainty $\pm 0.5\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.005\%$ Noise $\pm 0.00005\%$ $\pm 0.005\%$ Noise $\pm 0.00005\%$ $\pm 0.005\%$ Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ $\%$ Total Uncertainty $\pm 5.27\%$ $\pm 5.54\%$ Hereucy = 1 GHz; Sourcewith 80301AThermocouple Meter/SensorInstrumentation Uncertainty $\pm 0.925\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0.\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0.23\%$ $\pm 0.23\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.005\%$ $\pm 5\%$ Noise $\pm 0.005\%$ $\pm 5\%$ Noise $\pm 0.005\%$ $\pm 10\%$ Calibrator Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.005\%$	$\begin{array}{c c c c c c c c c c c c c c c c c c c $	dB Total Uncertainty	± 0.41 dB	+ 0.316 – 0.4 dB
OrderHyperterMatch = 1.5:1Sourcewith 80301AThermocouple Meter/SensorInstrumentation Uncertainty $\pm 0.5\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.005\%$ Moise $\pm 0.000005\%$ $\pm 0.005\%$ Noise $\pm 0.000005\%$ $\pm 0.005\%$ Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ $\%$ Total Uncertainty $\pm 5.27\%$ $\pm 5.54\%$ dB Total Uncertainty ± 0.22 dB ± 0.23 dB-30 dBm8650A Mother = 1.5:1Typical Meter/SensorInstrumentation Uncertainty $\pm 0.925\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0.28\%$ $\pm 0.23\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.005\%$ $\pm 5\%$ Noise $\pm 0.005\%$ $\pm 5\%$ Noise $\pm 0.005\%$ $\pm 0.005\%$	OutputOutputMatch = 1.5:100001ThermocoupleInstrumentation Uncertainty $\pm 0.5\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Vsensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.00005\%$ Noise $\pm 0.000005\%$ $\pm 0.005\%$ Noise $\pm 0.000005\%$ $\pm 0.005\%$ Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ $\%$ Total Uncertainty $\pm 5.27\%$ $\pm 5.54\%$ dB Total Uncertainty ± 0.22 dB ± 0.23 dB-30 dBm8650ATypicalFrequency = 1 GHz; SourcewithThermocoupleMatch = 1.5:180301AMeter/SensorInstrumentation Uncertainty $\pm 0.925\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ Noise $\pm 0.005\%$ $\pm 5\%$ Noise $\pm 0.005\%$ $\pm 5\%$ Noise $\pm 0.005\%$ $\pm 10\%$ Mismatch (Sensor/So	0 dBm	8650.0	Typical
Match = 1.5:180301AMeter/SensorInstrumentation Uncertainty $\pm 0.5\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibration Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.00005\%$ $\pm 0.005\%$ Noise $\pm 0.00005\%$ $\pm 0.01\%$ Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ $\%$ Total Uncertainty $\pm 5.27\%$ $\pm 5.54\%$ dB Total Uncertainty ± 0.22 dB ± 0.23 dB-30 dBm8650ATypicalFrequency = 1 GHz; Sourcewith 80301AMeter/SensorInstrumentation Uncertainty $\pm 0.925\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0.28\%$ $\pm 0.23\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibration Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.005\%$ $\pm 5\%$ Noise $\pm 0.005\%$ $\pm 5\%$ Noise $\pm 0.005\%$ $\pm 5\%$ Noise $\pm 0.005\%$ $\pm 10\%$	Match = 1.5:1 80301A Meter/Sensor Instrumentation Uncertainty \pm 0.5% \pm 0.5% Sensor Power Linearity (>8 GHz) \pm 0% \pm 0% Calibrator Uncertainty \pm 1.2% \pm 1.2% Calibrator Uncertainty \pm 1.2% \pm 1.2% Calibrator/Sensor Mismatch \pm 0.28% \pm 0.23% Calibrator Factor Uncertainty \pm 1.04% \pm 1.6% Zero Set \pm 0.000005% \pm 0.005% Noise \pm 0.000005% \pm 0.01% Mismatch (Sensor/Source) \pm 2.25% \pm 2.0% $\%$ Total Uncertainty \pm 5.27% \pm 5.54% dB Total Uncertainty \pm 0.22 dB \pm 0.23 dB -30 dBm 8650A Typical Frequency = 1 GHz; Source with Thermocouple Match = 1.5:1 80301A Meter/Sensor Instrumentation Uncertainty \pm 0.925% \pm 0.5% Sensor Power Linearity (>8 GHz) \pm 0% \pm 0% Calibrator Uncertainty \pm 1.2% \pm 1.2% Calibrator Uncertainty	Frequency = 1 GHz: Source	with	Thermocouple
Instrumentation Uncertainty $\pm 0.5\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibration Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.00005\%$ $\pm 0.005\%$ Noise $\pm 0.00005\%$ $\pm 0.001\%$ Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ $\%$ Total Uncertainty $\pm 5.27\%$ $\pm 5.54\%$ dB Total Uncertainty ± 0.22 dB ± 0.23 dB-30 dBm8650ATypicalFrequency = 1 GHz; Sourcewith 80301AMeter/SensorInstrumentation Uncertainty $\pm 0.925\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0.28\%$ $\pm 0.23\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.6\%$ Zero Set $\pm 0.005\%$ $\pm 5\%$ Noise $\pm 0.005\%$ $\pm 5\%$ Noise $\pm 0.005\%$ $\pm 5\%$ Noise $\pm 0.005\%$ $\pm 0.005\%$	$\begin{tabular}{ c c c c c c c c c c c c c c c c c c c$	Match = 1.5:1	80301A	Meter/Sensor
Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibration Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.00005\%$ $\pm 0.005\%$ Noise $\pm 0.00005\%$ $\pm 0.005\%$ Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ $\%$ Total Uncertainty $\pm 5.27\%$ $\pm 5.54\%$ dB Total Uncertainty ± 0.22 dB ± 0.23 dB-30 dBm8650ATypicalFrequency = 1 GHz; Sourcewith 80301AMeter/SensorInstrumentation Uncertainty $\pm 0.925\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0.28\%$ $\pm 0.23\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.005\%$ $\pm 5\%$ Noise $\pm 0.005\%$ $\pm 5\%$ Noise $\pm 0.005\%$ $\pm 0.005\%$	$\begin{array}{rrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrr$	Instrumentation Uncertainty	± 0.5%	± 0.5%
Calibrator Uncertainty \pm 1.2% \pm 1.2% Calibrator/Sensor Mismatch \pm 0.28% \pm 0.23% Calibration Factor Uncertainty \pm 1.04% \pm 1.6% Zero Set \pm 0.00005% \pm 0.005% Noise \pm 0.00005% \pm 0.01% Mismatch (Sensor/Source) \pm 2.25% \pm 2.0% % Total Uncertainty \pm 5.27% \pm 5.54% dB Total Uncertainty \pm 0.22 dB \pm 0.23 dB -30 dBm 8650A Typical Frequency = 1 GHz; Source with Thermocouple Match = 1.5:1 80301A Meter/Sensor Instrumentation Uncertainty \pm 0.925% \pm 0.5% Sensor Power Linearity (>8 GHz) \pm 0% \pm 0% Calibrator Uncertainty \pm 1.2% \pm 1.2% Calibrator/Sensor Mismatch \pm 0.28% \pm 0.23% Calibrator Factor Uncertainty \pm 1.04% \pm 1.6% Zero Set \pm 0.005% \pm 5% Noise \pm 0.005% \pm 5% Noise \pm 0.005% \pm 0%	$\begin{array}{rrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrr$	Sensor Power Linearity (>8 GHz)	± 0%	± 0%
Calibrator/Sensor Mismatch \pm 0.28% \pm 0.23% Calibration Factor Uncertainty \pm 1.04% \pm 1.6% Zero Set \pm 0.00005% \pm 0.005% Noise \pm 0.00005% \pm 0.01% Mismatch (Sensor/Source) \pm 2.25% \pm 2.0% % Total Uncertainty \pm 5.27% \pm 5.54% dB Total Uncertainty \pm 0.22 dB \pm 0.23 dB -30 dBm 8650A Typical Frequency = 1 GHz; Source with Thermocouple Match = 1.5:1 80301A Meter/Sensor Instrumentation Uncertainty \pm 0.925% \pm 0.5% Sensor Power Linearity (>8 GHz) \pm 0% \pm 0% Calibrator Uncertainty \pm 1.2% \pm 1.2% Calibrator Sensor Mismatch \pm 0.28% \pm 0.23% Calibration Factor Uncertainty \pm 1.04% \pm 1.6% Zero Set \pm 0.005% \pm 5% Noise \pm 0.005% \pm 0%	$\begin{array}{rrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrr$	Calibrator Uncertainty	± 1.2%	± 1.2%
Calibration Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.00005\%$ $\pm 0.005\%$ Noise $\pm 0.00005\%$ $\pm 0.005\%$ Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ % Total Uncertainty $\pm 5.27\%$ $\pm 5.54\%$ dB Total Uncertainty ± 0.22 dB ± 0.23 dB -30 dBm 8650A Typical Frequency = 1 GHz; Source with Thermocouple Match = 1.5:1 80301A Meter/Sensor Instrumentation Uncertainty $\pm 0.925\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0.28\%$ $\pm 0.23\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibration Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.005\%$ $\pm 5\%$ Noise $\pm 0.005\%$ $\pm 5\%$	$\begin{array}{r c c c c c c c c c c c c c c c c c c c$	Calibrator/Sensor Mismatch	± 0.28%	± 0.23%
Zero Set $\pm 0.00005\%$ $\pm 0.005\%$ Noise $\pm 0.00005\%$ $\pm 0.01\%$ Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ % Total Uncertainty $\pm 5.27\%$ $\pm 5.54\%$ dB Total Uncertainty ± 0.22 dB ± 0.23 dB -30 dBm 8650A Typical Frequency = 1 GHz; Source with Thermocouple Match = 1.5:1 80301A Meter/Sensor Instrumentation Uncertainty $\pm 0.925\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibration Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.005\%$ $\pm 5\%$ Noise $\pm 0.005\%$ $\pm 5\%$	$\begin{array}{rrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrrr$	Calibration Factor Uncertainty	± 1.04%	± 1.6%
Noise ± 0.00005% ± 0.01% Mismatch (Sensor/Source) ± 2.25% ± 2.0% % Total Uncertainty ± 5.27% ± 5.54% dB Total Uncertainty ± 0.22 dB ± 0.23 dB -30 dBm 8650A Typical Frequency = 1 GHz; Source with Thermocouple Match = 1.5:1 80301A Meter/Sensor Instrumentation Uncertainty ± 0.925% ± 0.5% Sensor Power Linearity (>8 GHz) ± 0% ± 0% Calibrator Uncertainty ± 1.2% ± 1.2% Calibrator/Sensor Mismatch ± 0.28% ± 0.23% Calibration Factor Uncertainty ± 1.04% ± 1.6% Zero Set ± 0.005% ± 5% Noise ± 0.005% ± 0%	$\begin{tabular}{ c c c c c }\hline Noise & \pm 0.00005\% & \pm 0.01\% \\\hline Mismatch (Sensor/Source) & \pm 2.25\% & \pm 2.0\% \\\hline \hline Mismatch (Sensor/Source) & \pm 2.25\% & \pm 2.0\% \\\hline \hline & \hline & Total Uncertainty & \pm 5.27\% & \pm 5.54\% \\\hline & dB Total Uncertainty & \pm 0.22 \ dB & \pm 0.23 \ dB \\\hline & Frequency = 1 \ GHz; \ Source & with & Thermocouple \\\hline Match = 1.5:1 & 80301A & Meter/Sensor \\\hline & Instrumentation Uncertainty & \pm 0.925\% & \pm 0.5\% \\\hline & Sensor Power Linearity (>8 \ GHz) & \pm 0\% & \pm 0\% \\\hline & Calibrator Uncertainty & \pm 1.2\% & \pm 1.2\% \\\hline & Calibrator Uncertainty & \pm 1.2\% & \pm 0.23\% \\\hline & Calibrator Sensor Mismatch & \pm 0.28\% & \pm 0.23\% \\\hline & Calibrator Sensor Mismatch & \pm 0.005\% & \pm 5\% \\\hline & Noise & \pm 0.005\% & \pm 10\% \\\hline & Noise & \pm 0.005\% & \pm 10\% \\\hline & Mismatch (Sensor/Source) & \pm 2.25\% & \pm 2.0\% \\\hline & \% \ Total Uncertainty & \pm 5.71\% & \pm 20.53\% \\\hline & dB \ Total Uncertainty & \pm 0.24 \ dB & \pm 0.8 \ dB \\\hline \end{tabular}$	Zero Set	± 0.000005%	± 0.005%
Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ % Total Uncertainty $\pm 5.27\%$ $\pm 5.54\%$ dB Total Uncertainty ± 0.22 dB ± 0.23 dB-30 dBm8650ATypicalFrequency = 1 GHz; SourcewithThermocoupleMatch = 1.5:180301AMeter/SensorInstrumentation Uncertainty $\pm 0.925\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibration Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.005\%$ $\pm 5\%$ Noise $\pm 0.005\%$ $\pm 10\%$	$\begin{array}{rrrr} \hline \mbox{Mismatch (Sensor/Source)} & \pm 2.25\% & \pm 2.0\% \\ \hline \mbox{Total Uncertainty} & \pm 5.27\% & \pm 5.54\% \\ \hline \mbox{dB Total Uncertainty} & \pm 0.22 \ dB & \pm 0.23 \ dB \\ \hline \mbox{Frequency} = 1 \ GHz; \ \mbox{Source} & with \\ \hline \mbox{Thermocouple} \\ \hline \mbox{Match} = 1.5:1 & 80301A & \mbox{Meter/Sensor} \\ \hline \mbox{Instrumentation Uncertainty} & \pm 0.925\% & \pm 0.5\% \\ \hline \mbox{Sensor Power Linearity} & \pm 0.925\% & \pm 0.0\% \\ \hline \mbox{Calibrator Uncertainty} & \pm 1.2\% & \pm 1.2\% \\ \hline \mbox{Calibrator/Sensor Mismatch} & \pm 0.28\% & \pm 0.23\% \\ \hline \mbox{Calibrator/Sensor Mismatch} & \pm 0.005\% & \pm 5\% \\ \hline \mbox{Noise} & \pm 0.005\% & \pm 10\% \\ \hline \mbox{Mismatch (Sensor/Source)} & \pm 2.25\% & \pm 2.0\% \\ \hline \mbox{Mismatch (Sensor/Source)} & \pm 2.25\% & \pm 2.0\% \\ \hline \mbox{M Total Uncertainty} & \pm 0.71\% & \pm 20.53\% \\ \hline \mbox{dB Total Uncertainty} & \pm 0.24 \ dB & \pm 0.8 \ dB \\ \hline \end{tabular}$	Noise	± 0.000005%	± 0.01%
$\%$ Total Uncertainty \pm 5.27% \pm 5.54%dB Total Uncertainty \pm 0.22 dB \pm 0.23 dB-30 dBm8650ATypicalFrequency = 1 GHz; SourcewithThermocoupleMatch = 1.5:180301AMeter/SensorInstrumentation Uncertainty \pm 0.925% \pm 0.5%Sensor Power Linearity (>8 GHz) \pm 0% \pm 0%Calibrator Uncertainty \pm 1.2% \pm 1.2%Calibrator/Sensor Mismatch \pm 0.28% \pm 0.23%Calibration Factor Uncertainty \pm 1.04% \pm 1.6%Zero Set \pm 0.005% \pm 5%Noise \pm 0.005% \pm 10%	$\label{eq:spectral_stress} \hline \begin{array}{ c c c c } & & & & & & & & & & & & & & & & & & &$	Mismatch (Sensor/Source)	± 2.25%	± 2.0%
dB Total Uncertainty \pm 0.22 dB \pm 0.23 dB-30 dBm8650ATypicalFrequency = 1 GHz; SourcewithThermocoupleMatch = 1.5:180301AMeter/SensorInstrumentation Uncertainty \pm 0.925% \pm 0.5%Sensor Power Linearity (>8 GHz) \pm 0% \pm 0%Calibrator Uncertainty \pm 1.2% \pm 1.2%Calibrator/Sensor Mismatch \pm 0.28% \pm 0.23%Calibration Factor Uncertainty \pm 1.04% \pm 1.6%Zero Set \pm 0.005% \pm 5%Noise \pm 0.005% \pm 10%	$\begin{tabular}{ c c c c } & dB \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ $	% Total Uncertainty	± 5.27%	± 5.54%
-30 dBm8650A withTypical Thermocouple Match = 1.5:1Instrumentation Uncertainty $\pm 0.925\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibrator/Sensor Mismatch $\pm 0.005\%$ $\pm 5\%$ Noise $\pm 0.005\%$ $\pm 5\%$	$\begin{tabular}{ c c c c c }\hline & & & & & & & & & & & & & & & & & & &$	dB Total Uncertainty	± 0.22 dB	± 0.23 dB
-30 dBm 8650A Typical Frequency = 1 GHz; Source with Thermocouple Match = 1.5:1 80301A Meter/Sensor Instrumentation Uncertainty ± 0.925% ± 0.5% Sensor Power Linearity (>8 GHz) ± 0% ± 0% Calibrator Uncertainty ± 1.2% ± 1.2% Calibrator/Sensor Mismatch ± 0.28% ± 0.23% Calibration Factor Uncertainty ± 1.04% ± 1.6% Zero Set ± 0.005% ± 5% Noise ± 0.005% ± 10%	$\begin{tabular}{ c c c c c } \hline -30 \ dBm & 8650A & Typical \\ \hline Frequency = 1 \ GHz; \ Source & with & Thermocouple \\ \hline Match = 1.5:1 & 80301A & Meter/Sensor \\ \hline Instrumentation Uncertainty & \pm 0.925\% & \pm 0.5\% \\ \hline Sensor Power Linearity (>8 \ GHz) & \pm 0\% & \pm 0\% \\ \hline Calibrator Uncertainty & \pm 1.2\% & \pm 1.2\% \\ \hline Calibrator Sensor Mismatch & \pm 0.28\% & \pm 0.23\% \\ \hline Calibrator Incertainty & \pm 1.04\% & \pm 1.6\% \\ \hline Zero \ Set & \pm 0.005\% & \pm 5\% \\ \hline Noise & \pm 0.005\% & \pm 10\% \\ \hline Mismatch (Sensor/Source) & \pm 2.25\% & \pm 2.0\% \\ \hline \% \ Total Uncertainty & \pm 5.71\% & \pm 20.53\% \\ \hline dB \ Total Uncertainty & \pm 0.24 \ dB & \pm 0.8 \ dB \\ \hline \end{tabular}$			
Frequency = 1 GHz; Source with Thermocouple Match = 1.5:1Thermocouple Moter/SensorInstrumentation Uncertainty $\pm 0.925\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibration Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.005\%$ $\pm 5\%$ Noise $\pm 0.005\%$ $\pm 10\%$	$\label{eq:response} \begin{array}{ c c c c } \hline Frequency = 1 GHz; Source with Meter/Sensor \\ \hline Match = 1.5:1 & 80301A & Meter/Sensor \\ \hline Instrumentation Uncertainty \pm 0.925\% \pm 0.5\% \\ \hline Sensor Power Linearity (>8 GHz) \pm 0\% \pm 0\% \\ \hline Calibrator Uncertainty \pm 1.2\% \pm 1.2\% \\ \hline Calibrator/Sensor Mismatch \pm 0.28\% \pm 0.23\% \\ \hline Calibrator/Sensor Mismatch \pm 0.28\% \pm 0.23\% \\ \hline Calibrator Incertainty \pm 1.04\% \pm 1.6\% \\ \hline Zero Set \pm 0.005\% \pm 5\% \\ \hline Noise \pm 0.005\% \pm 10\% \\ \hline Mismatch (Sensor/Source) \pm 2.25\% \pm 2.0\% \\ \hline \% \ Total Uncertainty \pm 5.71\% \pm 20.53\% \\ \hline dB \ Total Uncertainty \pm 0.24 \ dB \ \pm 0.8 \ dB \\ \hline \end{array}$	–30 dBm	8650A	Typical
Match = 1.5:180301AMeter/SensorInstrumentation Uncertainty $\pm 0.925\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibration Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.005\%$ $\pm 5\%$ Noise $\pm 0.005\%$ $\pm 10\%$	Match = 1.5:180301AMeter/SensorInstrumentation Uncertainty $\pm 0.925\%$ $\pm 0.5\%$ Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibration Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.005\%$ $\pm 5\%$ Noise $\pm 0.005\%$ $\pm 10\%$ Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$ % Total Uncertainty $\pm 5.71\%$ $\pm 20.53\%$ dB Total Uncertainty ± 0.24 dB ± 0.8 dB	Frequency = 1 GHz; Source	with	Thermocouple
Instrumentation Uncertainty ± 0.925% ± 0.5% Sensor Power Linearity (>8 GHz) ± 0% ± 0% Calibrator Uncertainty ± 1.2% ± 1.2% Calibrator/Sensor Mismatch ± 0.28% ± 0.23% Calibration Factor Uncertainty ± 1.04% ± 1.6% Zero Set ± 0.005% ± 5% Noise ± 0.005% ± 10%	Instrumentation Uncertainty ± 0.925% ± 0.5% Sensor Power Linearity (>8 GHz) ± 0% ± 0% Calibrator Uncertainty ± 1.2% ± 1.2% Calibrator/Sensor Mismatch ± 0.28% ± 0.23% Calibration Factor Uncertainty ± 1.04% ± 1.6% Zero Set ± 0.005% ± 5% Noise ± 0.005% ± 10% Mismatch (Sensor/Source) ± 2.25% ± 2.0% % Total Uncertainty ± 5.71% ± 20.53% dB Total Uncertainty ± 0.24 dB ± 0.8 dB	Match = 1.5:1	80301A	Meter/Sensor
Sensor Power Linearity (>8 GHz) $\pm 0\%$ $\pm 0\%$ Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibration Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.005\%$ $\pm 5\%$ Noise $\pm 0.005\%$ $\pm 10\%$	$\begin{array}{llllllllllllllllllllllllllllllllllll$	Instrumentation Uncertainty	± 0.925%	± 0.5%
Calibrator Uncertainty $\pm 1.2\%$ $\pm 1.2\%$ Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibration Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.005\%$ $\pm 5\%$ Noise $\pm 0.005\%$ $\pm 10\%$	$\begin{array}{llllllllllllllllllllllllllllllllllll$	Sensor Power Linearity (>8 GHz)	± 0%	± 0%
Calibrator/Sensor Mismatch $\pm 0.28\%$ $\pm 0.23\%$ Calibration Factor Uncertainty $\pm 1.04\%$ $\pm 1.6\%$ Zero Set $\pm 0.005\%$ $\pm 5\%$ Noise $\pm 0.005\%$ $\pm 10\%$	$\begin{array}{c cl} Calibrator/Sensor Mismatch & \pm 0.28\% & \pm 0.23\% \\ \hline Calibration Factor Uncertainty & \pm 1.04\% & \pm 1.6\% \\ \hline Zero Set & \pm 0.005\% & \pm 5\% \\ \hline Noise & \pm 0.005\% & \pm 10\% \\ \hline Mismatch (Sensor/Source) & \pm 2.25\% & \pm 2.0\% \\ \hline \% \mbox{ Total Uncertainty } & \pm 5.71\% & \pm 20.53\% \\ \hline dB \mbox{ Total Uncertainty } & \pm 0.24 \mbox{ dB } & \pm 0.8 \mbox{ dB } \\ \hline \end{array}$	Calibrator Uncertainty	± 1.2%	± 1.2%
Calibration Factor Uncertainty ± 1.04% ± 1.6% Zero Set ± 0.005% ± 5% Noise ± 0.005% ± 10%	Calibration Factor Uncertainty ± 1.04% ± 1.6% Zero Set ± 0.005% ± 5% Noise ± 0.005% ± 10% Mismatch (Sensor/Source) ± 2.25% ± 2.0% % Total Uncertainty ± 5.71% ± 20.53% dB Total Uncertainty ± 0.24 dB ± 0.8 dB	Calibrator/Sensor Mismatch	± 0.28%	± 0.23%
Zero Set ± 0.005% ± 5% Noise ± 0.005% ± 10%	Zero Set ± 0.005% ± 5% Noise ± 0.005% ± 10% Mismatch (Sensor/Source) ± 2.25% ± 2.0% % Total Uncertainty ± 5.71% ± 20.53% dB Total Uncertainty ± 0.24 dB ± 0.8 dB	Calibration Factor Uncertainty	± 1.04%	± 1.6%
Noise ± 0.005% ± 10%	Noise ± 0.005% ± 10% Mismatch (Sensor/Source) ± 2.25% ± 2.0% % Total Uncertainty ± 5.71% ± 20.53% dB Total Uncertainty ± 0.24 dB ± 0.8 dB	Zero Set	± 0.005%	± 5%
N° 11 (C (C) 0.050(0.00(Mismatch (Sensor/Source) ± 2.25% ± 2.0% % Total Uncertainty ± 5.71% ± 20.53% dB Total Uncertainty ± 0.24 dB ± 0.8 dB	Noise	± 0.005%	± 10%
Mismatch (Sensor/Source) $\pm 2.25\%$ $\pm 2.0\%$	% Total Uncertainty ± 5.71% ± 20.53% dB Total Uncertainty ± 0.24 dB ± 0.8 dB	Mismatch (Sensor/Source)	± 2.25%	± 2.0%
% Total Uncertainty ± 5.71% ± 20.53%	dB Total Uncertainty ± 0.24 dB ± 0.8 dB	% Total Uncertainty	± 5.71%	± 20.53%
dB Total Uncertainty ± 0.24 dB ± 0.8 dB		dB Total Uncertainty	± 0.24 dB	± 0.8 dB

enter new calibration factors. You not only avoid measurement errors; you also save yourself test time.

An EEPROM in all Giga-tronics sensors automatically applies the correct cal factor, so you save time and avoid measurement errors.



Giga-tronics power meter architecture provides for a broad choice of functional sensors. Just by changing a sensor, you can measure CW power, pulse power, and the peak and average power of TDMA, GSM and CDMA signals faster, more accurately, and over a wider range.



THE FASTEST CW MEASUREMENTS

Giga-tronics 80300A Series CW Power Sensors let you measure CW power from 10 MHz to 40 GHz at more than 1,750 readings per second over GPIB.

Measure up to 90 dB with a single sensor, and select from a variety of high power sensors, up to 50 W.

PULSE POWER MEASUREMENTS

Attach a Giga-tronics 80350A Series Peak Power Sensor to an 8650A meter and directly measure the instantaneous peak power level of a pulse modulated signal.

Use the 'sample delay' function to set the desired measurement point on the waveform. And an external scope can be used to view the profile and see the exact measurement point on the pulse.



Sensor	Measurement	Canabilities
3011301	wicasurement	capabilities

			Sensor Mod	el	
Signal Type	80301A	80350A	80401A	80601A	80701A
CW Power Level	–70 to +20 dBm	-30 to +20 dBm	–67 to +20 dBm	–67 to +20 dBm	-64 to +20 dBm
Amplitude Modulation	N/A	N/A	$f_m \le 40 \text{ kHz}$, -60 to +20 dBm	$f_m \le 1.5 \text{ MHz}$, -60 to +20 dBm	$f_m \le 10 \text{ MHz}$, -60 to +20 dBm
Rate, Power Range			$f_m > 40 \text{ kHz}$, -60 to -20 dBm	$f_m > 1.5 \text{ MHz}$, -60 to -20 dBm	
Two-Tone	N/A	N/A	≤ 40 kHz, –60 to +20 dBm	≤ 1.5 MHz, –60 to +20 dBm	\leq 10 MHz, –60 to +20 dBm
Maximum Separation			> 40 kHz, -60 to -20 dBm	> 1.5 MHz, -60 to -20 dBm	> 10 MHz, -60 to -20 dBm
Between Carriers					
Pulse Modulation	N/A	> 350 ns	> 200 µs	> 300 µs	> 100 µs
		Pulse Width	Pulse Width	Pulse Width	Pulse Width
Burst with Modulation	N/A	N/A	$f_m \le 40 \text{ kHz}$, > 200 μs	$f_m \le 1.5 \text{ MHz}$, > 300 µs	$f_m \le 10 \text{ MHz}$, > 100 µs
f _{m = modulation rate}			Pulse Width; -40 to +20 dBm	Pulse Width; -40 to +20 dBm	Pulse Width; -30 to +20 dBm
			$f_m > 40 \text{ kHz}, > 200 \mu\text{s}$	f _m > 1.5 MHz, > 300 μs	f _m > 10 MHz, > 100 μs
			Pulse Width; -40 to -20 dBm	Pulse Width; -40 to -20 dBm	Pulse Width; -30 to -20 dBm

MODULATED POWER MEASUREMENTS

The Giga-tronics 80400A Series Modulated Power Sensors let you measure the average power of amplitude modulated, burst modulated and other complex modulated signals — such



as TDMA signals — at bandwidths up to 40 kHz.

The Giga-tronics 80600A Series Modulated Power Sensors provide bandwidth up to 1.5 MHz to measure the peak and average power of CDMA signals.



And the Giga-tronics 80701A Modulated Power Sensor operating with the 8650A power meter, provides system bandwidth up to 10 MHz to measure the peak and average power of wide band, third-generation CDMA signals over an 80 dB range.



SEE FOR YOURSELF

The 8650A incorporates a 3.72" wide by 2.15" high Liquid Crystal Display (LCD) with 240 x 120 dot resolution, 0.38 mm dot pitch, and Cold Cathode Fluorescent Lamp (CCFL) back light for maximum detail and optimum viewing.

The large display lets you see more information. And the display works in tandem with the meter controls to let you view menu selections and see your input data as you enter it.

You can view calibration information, select a standard mode, setup and recall preconfigured, custom modes, and set measurement points and durations.

An extensive list of help panels provide assistance in setting up special features and guidance in making the measurement.



Each sensor uses an EEPROM to store values of cal factor. Entering the measurement frequency





Sensor Freq Compensation

The graphic display provides visual feedback as you set the measurement start time and duration of the time gate to measure the average power during a specific time period.



View the mean power and standard deviation of the modulated signal over a time period of interest. Standard deviation offers an alternative descriptive analysis of the power variation when compared to the traditional crest factor.

The Cumulative Distribution Function (CDF) shows the percentage of time a signal is below a selected power level. The x axis displays the amount of power at the selected level, measured in dBm, and the y axis displays the percentage of time the power is at or below the power specified by the x axis. The Complementary Cumulative Distribution Function (CCDF) reorients the CDF curve in accordance with the equation CCDF = 1-CDFfor more accustomed viewing of a descending slope. Moving a cursor along the slope of the curve displays the power level in dBm and the corresponding percentage of time the signal is above that level.





The histogram function allows you to view a power range distribution over a period of time. The x axis displays the minimum to maximum power levels measured during the interval time period, and the y axis displays the percent of time each power level is measured. A zoom feature lets you view smaller segments of the power range to better analyze the percentage of time a specific power level has occurred.

The strip chart function allows you to view the vary-

ing power levels of a signal over a period of time. The x axis displays time from the start of the measurement to a selectable period of 1 to 200 minutes, and the y axis displays the minimum to maximum power levels measured during the selected period. Moving a cursor along the x axis displays time and the corresponding power level.



STATISTICAL ANALYSIS

Excessive cost can prove as detrimental to the success of communications equipment as inadequate performance.

The 8650A provides a range of statistical power measurement analysis features that help you optimize your designs to prevent inadequate performance due to under design or excessive cost due to over design.

These features include crest factor, standard deviation, strip chart, CDF/CCDF, and histogram, and they let you view and thoroughly analyze the power signal over a selected period of time.

Combined, they make the 8650A the most advanced power meter available for communications systems design.

Giga-li O	The sense of the sense	Selection Guide						
	Frequency Range/ Power Range	Maximum Power	Power Linearity 4 (Frequency > 8 GHz)	RF Connector	Length	Diameter	Weight	VSWR
200 mW	CW Power Sensors							
80301A	10 MHz to 18 GHz -70 to +20 dBm	+23 dBm (200 mW)	_70 to _20 dBm: ±0.00 dB _20 to +20 dBm: ±0.05 dB/10 dB	Type N(m) 50Ω	114.5 mm (4.5 in)	32 mm (1.25 in)	0.18 kg (0.4 lb)	1.12: 0.01 - 2 GHz 1.22: 2 - 12.4 GHz
80302A	10 MHz to 18 GHz _70 to +20 dBm	+23 dBm (200 mW)	–70 to –20 dBm: ±0.00 dB –20 to +20 dBm: ±0.05 dB/10 dB	APC-7 50Ω	114.5 mm (4.5 in)	32 mm (1.25 in)	0.18 kg (0.4 lb)	1.29: 12.4 - 18 GHz
80303A	10 MHz to 26.5 GHz 	+23 dBm (200 mW)	–70 to –20 dBm: ±0.00 dB –20 to +20 dBm: ±0.1 dB/10 dB	Type K(m) 1 50Ω	114.5 mm (4.5 in)	32 mm (1.25 in)	0.18 kg (0.4 lb)	1.12: 0.01 - 2 GHz 1.22: 2 - 12.4 GHz
80304A	10 MHz to 40 GHz -70 to 0 dBm	+23 dBm (200 mW)	-70 to -20 dBm: ±0.00 dB -20 to 0 dBm: ±0.2 dB/10 dB	Type K(m) ¹ 50Ω	114.5 mm (4.5 in)	32 mm (1.25 in)	0.18 kg (0.4 lb)	1.38: 12.4 - 18 GHz 1.43: 18 - 26.5 GHz 1.92: 26.5 - 40 GHz
Low VS	WR CW Power Sensors							
80310A	10 MHz to 18 GHz 64 to +26 dBm	+29 dBm (800 mW)	_64 to _14 dBm: ±0.00 dB _14 to +26 dBm: ±0.05 dB/10 dB	Type K(m) ¹ 50Ω	127 mm (5.0 in)	32 mm (1.25 in)	0.23 kg (0.5 lb)	1.13: 0.01 - 2 GHz 1.16: 2 - 12 GHz
80313A	10 MHz to 26.5 GHz 64 to +26 dBm	+29 dBm (800 mW)	_64 to _14 dBm: ±0.00 dB _14 to +26 dBm: ±0.1 dB/10 dB					1.23: 12 - 18 GHz 1.29: 18 - 26.5 GHz
80314A	10 MHz to 40 GHz 64 to +6 dBm	+29 dBm (800 mW)	-64 to -14 dBm: ±0.00 dB -14 to +6dBm: ±0.2 dB/10 dB					1.50: 26.5 - 40 GHz
1 W CW	Power Sensors							
80320A	10 MHz to 18 GHz _60 to +30 dBm	+30 dBm (1 W)	-60 to −10 dBm:±0.00 dB −10 to +30 dBm: ±0.05 dB/10 dB	Type K(m) ¹ 50Ω	127 mm (5.0 in)	32 mm (1.25 in)	0.23 kg (0.5 lb)	1.11: 0.01 - 2 GHz 1.12: 2 - 12 GHz
80323A	10 MHz to 26.5 GHz 60 to +30 dBm	+30 dBm (1 W)	_60 to _10 dBm: ±0.00 dB _10 to +30 dBm: ±0.1 dB/10 dB					1.18: 12 - 18 GHz 1.22: 18 - 26.5 GHz
80324A	10 MHz to 40 GHz _60 to +10 dBm	+30 dBm (1 W)	_60 to _10 dBm: ±0.00 dB _10 to +10 dBm: ±0.2 dB/10 dB					1.36: 26.5 - 40 GHz
<u>5 W CW</u>	Power Sensor ²							
80321A	10 MHz to 18 GHz -50 to +37 dBm	+37 dBm (5 W)	_50 to 0 dBm: ±0.00 dB 0 to +37 dBm: ±0.05 dB/10 dB	Type N(m) 50Ω	150 mm (5.9 in)	32 mm (1.25 in)	0.23 kg (0.5 lb)	1.20: 0.01 - 6 GHz 1.25: 6 - 12.4 GHz 1.35: 12.4 - 18 GHz
25 W CV	V Power Sensor ³							
80322A	10 MHz to 18 GHz -40 to +44 dBm	+44 dBm (25 W)	-40 to +10 dBm: ±0.00 dB +10 to +44 dBm: ±0.05 dB/10 dB	Type N(m) 50Ω	230 mm (9.0 in)	104 mm (4.1 in)	0.3 kg (0.6 lb)	1.20: 0.01 - 6 GHz 1.30: 6 - 12.4 GHz 1.40: 12.4 - 18 GHz
50 W CV	V Power Sensor ³							
80325A	10 MHz to 18 GHz _40 to +47 dBm	+47 dBm (50 W)	-40 to +10 dBm: ±0.00 dB +10 to +47 dBm: ±0.05 dB/10 dB	Type N(m) 50Ω	230 mm (9.0 in)	104 mm (4.1 in)	0.3 kg (0.6 lb)	1.25: 0.01 - 6 GHz 1.35: 6 - 12.4 GHz 1.45: 12.4 - 18 GHz

Giga-tro	nics Peak Power Senso	r Selection Guide						
	Frequency Range/ Power Range	Maximum Power	Power Linearity ⁴ (Frequency > 8 GHz)	RF Connector	Length	Diameter	Weight	VSWR
200 mW	Peak Power Sensors							
80350A	45 MHz to 18 GHz	+23 dBm (200 mW)	_30 to _20 dBm: ±0.00 dB	Type N(m)	165 mm	37 mm	0.3 kg	1.12: 0.045 - 2 GHz
	_20 to +20 dBm, Peak	CW or Peak	_20 to +20 dBm: ±0.05 dB/10 dB	50Ω	(6.5 in)	(1.25 in)	(0.7 lb)	1.22: 2 - 12.4 GHz
	<u>-30 to +20 dBm, CW</u>							1.37: 12.4 - 18 GHz
80353A	45 MHz to 26.5 GHz	+23 dBm (200 mW)	-30 to -20 dBm: ±0.00 dB	Type K(m)	165 mm	37 mm	0.3 kg	1.50: 18 - 26.5 GHz
	_20 to +20 dBm, Peak	CW or Peak	=20 to +20 dBm: ±0.1 dB/10 dB	50Ω	(6.5 IN)	(1.25 IN)	(di 1.0)	1.92: 26.5 - 40 GHZ
002544		. 22 dDm (200 m)///	20 to _ 20 dBm; _ 0.00 dB	Turne K(m) 1	1/E mm	27 mm	0.2 kg	
80354A	40 IVIAZ LO 40 GAZ	+23 UBIII (200 IIIVV)	-30 to 0.0 dBm; +0.2 dB/10 dB	Type K(III)	(6 5 in)	37 [[][] (1.25 in)	0.3 Kg (0.7 lb)	
	20 to +0.0 dBm CW	CW UI FEAK	=20 to 0.0 dbin. ±0.2 db/ 10 db	5052	(0.5 11)	(1.25 11)	(0.7 10)	
5 W Pea	k Power Sensor 5,7							
80351A	45 MHz to 18 GHz	CW: +37 dBm	_10 to +0 dBm: +0.00 dB	Type N(m)	200 mm	37 mm	0.3 kg	1.15: 0.045 - 4 GHz
	0 to +40 dBm. Peak	(5 W Average)	0.0 to +40 dBm: ±0.05 dB/10 dB	50Ω	(7.9 in)	(1.25 in)	(0.7 lb)	1.25: 4 - 12.4 GHz
	-10 to +37 dBm, CW	Peak: +43 dBm			· /			1.35: 12.4 - 18 GHz
25 W Pe	ak Power Sensor 6.7							
80352A	45 MHz to 18 GHz	CW: +44 dBm	0.0 to +10 dBm: ±0.00 dB	Type N(m)	280 mm	104 mm	0.3 kg	1.20: 0.045 - 6 GHz
	+10 to +50 dBm, Peak	(25 W Average)	+10 to +50 dBm: ±0.05 dB/10 dB	50Ω	(11.0 in)	(4.1 in)	(0.7 lb)	1.30: 6 - 12.4 GHz
	0.0 to +44 dBm, CW	Peak: +53 dBm						1.40: 12.4 - 18 GHz
50 W Pe	ak Power Sensor 6.7							
80355A	45 MHz to 18 GHz	CW: +47 dBm	0.0 to +10 dBm: ±0.00 dB	Type N(m)	280 mm	104 mm	0.3 kg	1.25: 0.045 - 6 GHz
	+ 10 to +50 dBm, Peak	(50 VV Average)	+10 to +50 aBm: ±0.05 dB/10 dB	5022	(11.0 IN)	(4.1 IN)	(U.7 lb)	1.35: 6 - 12.4 GHZ
	0.0 to +47 dBm, CW	Peak: +53 dBm						1.45: 12.4 - 18 GHZ

Giga-tro	nics Bridge Selection Guid	e						
Precisio	Frequency Range/ Power Range n CW Return Loss Bridges	Maximum Power	Power Linearity 4 (Frequency > 8 GHz)	Input	Test Port	Directivity	Weight	VSWR
80501	10 MHz to 18 GHz -35 to +20 dBm	+27 dBm (0.5 W)	_35 to +10 dBm: ±0.1 dB +10 to +20 dBm: ±0.1 dB ±0.005 dB/dB	Type N(f) 50Ω	Type N(f) 50Ω	38 dB	0.340 kg	< 1.17: 0.01 - 8 GHz < 1.27: 8 - 18 GHz
80502	10 MHz to 18 GHz -35 to +20 dBm	+27 dBm (0.5 W)	_35 to +10 dBm: ±0.1 dB +10 to +20 dBm: ±0.1 dB ±0.005 dB/dB	Type N(f) 50Ω	APC-7(f) 50Ω	40 dB	0.340 kg	< 1.13: 0.01 - 8 GHz < 1.22: 8 - 18 GHz
80503	10 MHz to 26.5 GHz 35 to +20 dBm	+27 dBm (0.5 W)	_35 to +10 dBm: ±0.1 dB +10 to +20 dBm: ±0.1 dB ±0.005 dB/dB	SMA(f) 50Ω	SMA(f) 50Ω	35 dB	0.340 kg	< 1.22: 0.01 - 18 GHz < 1.27: 18 - 26.5 GHz
80504	10 MHz to 40 GHz 35 to +20 dBm	+27 dBm (0.5 W)	-35 to +10 dBm: ±0.1 dB +10 to +20 dBm: ±0.1 dB ±0.005 dB/dB	Type K(f) 50Ω	Type K(f) 50Ω	30 dB	0.198 kg	< 1.35: 0.01 - 26.5 GHz < 1.44: 26.5 - 40 GHz

Giga-tronics Modulation Power Sensor Selection Guide ($f_m \le 40$ kHz)

	Frequency Range/ Power Range	Maximum Power	Power Linearity ⁴ (Frequency > 8 GHz)	RF Connector	Length	Diameter	Weight	VSWR
200 mW	Modulation Power Sen	sors						
80401A	10 MHz to 18 GHz	+23 dBm (200 mW)	_67 to _20 dBm: ±0.00 dB	Type N(m)	114.5 mm	32 mm	0.18 kg	1.12: 0.01 - 2 GHz
	_67 to +20 dBm		-20 to +20 dBm: ±0.05 dB/10 dB	50Ω	(4.5 in)	(1.25 in)	(0.4 lb)	1.22: 2 - 12.4 GHz
80402A	10 MHz to 18 GHz	+23 dBm (200 mW)	-67 to -20 dBm: ±0.00 dB	APC-7				1.29: 12.4 - 18 GHz
	_67 to +20 dBm		_20 to +20 dBm: ±0.05 dB/10 dB	50Ω				
Low VS	WR Modulation Power S	Sensor						
80410A	10 MHz to 18 GHz	+29 dBm (800 mW)	-64 to -14 dBm: ±0.00 dB	Type K 1(m)	127 mm	32 mm	0.23 kg	1.13: 0.01 - 2 GHz
	_64 to +26 dBm		_14 to +26 dBm: ±0.05 dB/10 dB	50Ω	(5.0 in)	(1.25 in)	(0.5 lb)	1.16: 2 - 12 GHz
								1.23: 12 - 18 GHz
<u>1 W Mo</u>	dulation Power Sensor							
80420A	10 MHz to 18 GHz	+30 dBm (1 W)	_57 to_10 dBm: ±0.00 dB	Type K ¹(m)	127 mm	32 mm	0.23 kg	1.11: 0.01 - 2 GHz
	_57 to +30 dBm		-10 to +30 dBm: ±0.05 dB/10 dB	50Ω	(5.0 in)	(1.25 in)	(0.5 lb)	1.12: 2 - 12 GHz
								1.18: 12 - 18 GHz
5 W Mo	dulation Power Sensor	2						
80421A	10 MHz to 18 GHz	+37 dBm (5 W)	_47 to 0 dBm: ±0.00 dB	Type N(m)	150 mm	32 mm	0.23 kg	1.20: 0.01 - 6 GHz
	_47 to +37 dBm		0 to +37 dBm: ±0.05 dB/10 dB	50Ω	(5.9 in)	(1.25 in)	(0.5 lb)	1.25: 6 - 12.4 GHz
								1.35: 12.4 - 18 GHz
<u>25 W Mo</u>	odulation Power Sensor	. 3						
80422A	10 MHz to 18 GHz	+44 dBm (25 W)	_37 to +10 dBm: ±0.00 dB	Type N(m)	230 mm	104 mm	0.3 kg	1.20: 0.01 - 6 GHz
	_37 to +44 dBm		+10 to +44 dBm: ±0.05 dB/10 dB	50Ω	(9.0 in)	4.1 in)	(0.6 lb)	1.30: 6 - 12.4 GHz
								1.40: 12.4 - 18 GHz
<u>50 W Mo</u>	odulation Power Sensor	3						
80425A	10 MHz to 18 GHz	+47 dBm (50 W)	_34 to +10 dBm: ±0.00 dB	Type N(m)	230 mm	104 mm	0.3 kg	1.25: 0.01 - 6 GHz
	_34 to +47 dBm		+10 to +47 dBm: ±0.05 dB/10 dB	50Ω	(9.0 in)	(4.1 in)	(0.6 lb)	1.35: 6 - 12.4 GHz
								1.45: 12.4 - 18 GHz

Giga-tronic	s wooulation Power	Sensor Selection Gui	de ($T_m \leq 1.5$ IVIHZ)					
F	Frequency Range/ Power Range	Maximum Power	Power Linearity 4 (Frequency > 8 GHz)	RF Connector	Length	Diameter	Weight	VSWR
200 mW M	odulation Power Sense	ors						
80601A 1 -	10 MHz to 18 GHz -67 to +20 dBm, CW	+23 dBm (200 mW)	67 to20 dBm: ±0.00 dB 20 to +20 dBm: ±0.05 dB/10 dB	Type N(m) 50Ω	137 mm (5.39 in)	41 mm (1.62 in)	0.23 kg (0.5 lb)	1.12: 0.01 - 2 GHz 1.22: 2 - 12.4 GHz 1.29: 12.4 - 18 GHz
5 W Modul	lation Power Sensor 5	7						
80621A 1 -	10 MHz to 18 GHz _47 to +37 dBm	+37 dBm (5 W)	–47 to 0 dBm: ±0.00 dB 0 to +37 dBm: ±0.05 dB/10 dB	Type N(m) 50Ω	175 mm (6.90 in)	41 mm (1.62 in)	0.28 kg (0.6 lb)	1.20: 0.01 - 6 GHz 1.25: 6 - 12.4 GHz 1.35: 12.4 - 18 GHz

Giga-tronics Modulation Power Sensor Selection Guide (f_{\rm m} \leq 10 MHz)

A · · ·

Frequency Range/ Power Range	Maximum Power	Power Linearity ⁴	RF Connector	Length	Diameter	Weight	VSWR
200 mW Modulation Power Sen	ISOF						
80701A (Requires Option 12)		Frequency >8 GHz					
50 MHz to 18 GHz	+23 dBm (200 mW)	-60 to -20 dBm: ±0.00 dB	Type N(m)	120 mm	27 mm	0.10 kg	1.12: 0.01 - 2 GHz
_64 to +20 dBm, CW		_20 to +20 dBm: ±0.05 dB/10 dB	50Ω	(4.72 in)	(1.06 in)	(0.2 lb)	1.22: 2 - 12.4 GHz
250 MHz to 18 GHz		Frequency <500 MHz					1.29: 12.4 - 18 GHz
-60 to +20 dBm, Modul	ation	-60 to -20 dBm: ±0.00 dB					
		20 to +20 dBm [·] +0 05 dB/10 dB					

Giga-tronics True RMS Sensors Selection Guide (f _m > 1.5 MHz) Frequency Range/ Power Linearity 4 Power Range Power (Frequency > 8 GHz) True RMS Sensors (-30 dBm to +20 dBm) 80330A 10 MHz to 18 GHz +33 dBm (2 W) -30 to +20 dBm: ±0.00 dB 80333A 10 MHz to 26.5 GHz -30 to +20 dBm: ±0.00 dB -30 to +20 dBm: ±0.00 dB										
		Frequency Range/ Power Range	Maximum Power	Power Linearity ⁴ (Frequency > 8 GHz)	RF Connector	Length	Diameter	Weight	VSWR	
True RMS Sensors (-30 dBm to +20 dBm)										
	80330A	10 MHz to 18 GHz	+33 dBm (2 W)	_30 to +20 dBm: ±0.00 dB	Type K(m) 1	152 mm	32 mm	0.27 kg	1.12: 0.01 - 12 GHz	
	80333A	10 MHz to 26.5 GHz			50Ω	(6.0 in)	(1.25 in)	(0.6 lb)	1.15: 12 - 18 GHz	
	80334A	10 MHz to 40 GHz							1.18: 18 - 26.5 GHz	
									1.29: 26.5 - 40 GHz	

Sensor Calibration Factor Uncertainties

Frequence	cy (GHz)		Root Su	im of Squa	res (RSS)	Uncertaint	ies(%) °	
		80301A				80321A°		
		80302A				80322A°		
		80350A				80325A°		
		80401A	80303A	80310A	80320A	80421A ⁹		
		80402A	80304A	80313A	80323A	80422A°	80330A	80351A [°]
		80601A	80353A	80314A	80324A	80425A ⁹	80333A	80352A ⁹
Lower	Upper	80701A	80354A	80410A	80420A	80621A ⁹	80334A	80355A°
Min	1	1.04	1.64	1.58	1.58	4.54	1.58	4.92
1	2	1.20	1.73	1.73	1.73	4.67	1.73	5.04
2	4	1.33	1.93	1.91	1.91	4.89	1.90	7.09
4	6	1.41	2.03	2.02	2.01	5.01	2.01	7.17
6	8	1.52	2.08	2.07	2.06	5.12	2.06	7.25
8	12.4	1.92	2.55	2.54	2.53	5.56	2.53	7.56
12.4	18	2.11	2.83	2.80	2.79	5.89	2.78	12.37
18	26.5	_	3.63	3.68	3.62	_	3.59	
26.5	40		6.05	5.54	5.39	_	5.30	_

¹ The K connector is electrically and mechanically compatible with the APC-3.5 and SMA connectors. Note: Use a Type N(m) to SMA(f) adapter (part no. 29835) for calibration of power sensors with Type K(m) connectors.² Power coefficient equals <0.01 dB/Watt.³ Power coefficient equals <0.015 dB/Watt. 4 For frequencies above 8 GHz, add power linearity to system linearity.⁵ Power coefficient equals <0.01 dB/Watt. 7 Peak operating range above CW maximum range is limited to <10% duty cycle.⁸ Square root of the sum of the individual uncertainties squared (RSS).⁹ Cal Factor numbers allow for 3% repeatability when reconnecting an attenuator to a sensor and 3% for attenuator measurement uncertainty and mismatch of sensor/pad combination.

Specifications describe the instrument's warranted performance, and apply when using the 80300A, 80400A, 80600A, and 80700A Series Sensors.

METER

Frequency Range: 10 MHz to 40 GHz 10 Power Range: -70 dBm to +47 dBm (100 pW to 50 Watt) 10 Single Sensor Dynamic Range: 10 CW Power Sensors: 90 dB Peak (Pulse) Power Sensors: 40 dB, Peak 50 dB, CW Modulation Power Sensors: 87 dB, CW 80 dB, MAP/PAP 11 60 dB, BAP 11 Display Resolution: User selectable from 1 dB to 0.001 dB in Log mode, and from 1 to 4 digits of display resolution in Linear mode. Meter Functions Measurement Modes (Sensors): CW (80300A, 80350A, 80400A, 80600A, and 80700A Series) Peak (80350A Series) MAP/PAP/BAP ¹¹ (80400A, 80600A, and 80700A Series) Averaging: User selectable, auto-averaging or manual from 1-512 readings. Timed averaging from 20 ms to 20 seconds. dB Rel and Offset: Power display can be offset by -99.999 to +99.999 dB to account for external loss/gain. **Configuration Storage Registers:** Allows up to 20 front panel setups. Power Measurements and Display Configurations: Any two of the following channel configurations, simultaneously: A, B, A/B, B/A, A-B, B-A, DLYA, DLYB Number of Display Lines: 4 Sampling: CW and Modulation Mode: 2.5 to 5 MHz asynchronous Analog Bandwidth: CW Mode: ≥3 kHz Modulation Mode: >10 MHz Time Gating: Trigger Delay: 0 to 327 ms Gate Time: 10 µs to 327 ms Holdoff Time: 0 to 327 ms ACCURACY 50 MHz Calibrator: (Standard)

 Calibrator: +20 dBm to -30 dBm power sweep calibration signal to dynamically linearize the power sensors.

 Connector: Type N, 50 Ω

 Frequency: 50 MHz, nominal

 0.0 dBm Accuracy: ±1.2% worst case for one year, over temperature range of 5° to 35°C.

 VSWR: <1.05 (Return Loss >33 dB) @ 0 dBm.





Input, (dBm)

Graph shows linearity plus worst case zero set, and noise versus input power

Temperature Coefficient of

Linearity: <0.3%/°C temperature change following Power Sweep calibration. 24 hour warm-up required. Zeroing Accuracy: (CW)

Zero Set: ¹² < \pm 50 pW, < \pm 100 pW with 80400A and 80600A Series Modulation Power Sensors. < \pm 200 pW with 80700A Series Sensors. Zero Drift: ¹² < \pm 100 pW during 1 hour, < \pm 200 pW with 80400A and 80600A Series Sensors, < \pm 400 pW with 80700A Series Sensors. Noise: < \pm 50 pW, < \pm 100 pW with 80400A and 80600A Series Modulation Power Sensors. < \pm 200 pW with 80700A Series Sensors. Measurable over any 1 minute interval after zeroing, 3 standard deviations.



- V Prop F Input (BNC): Sets calibration factors using source VpropF output.¹³
 - Analog Output (2) (BNC): Provides an output voltage of 0 to 10V for Channels 1 and 2 in either Lin or Log units.¹³ Does not operate in Swift or Buffered modes.

Trigger Input (BNC): TTL trigger input signal for Swift and Fast Buffered modes.

GPIB Interface: IEEE-488 and IEC-625 remote programming

RS232 Interface: Programmable serial interface, DB-9 connector

GENERAL SPECIFICATIONS

Temperature Range:

Operating: 0° to 55°C (+32° to +131°F)¹⁴ Storage: -40°C to 70°C (-40° to +158°F) Power Requirements: 100/120/220/240V ±10%, 48 to 440 Hz, 25VA typical

Physical Characteristics: Dimensions: 215 mm (8.4 in) wide,

89 mm (3.5 in) high, 368 mm (14.5 in) deep Weight: 4.55 kg (10lbs)

ORDERING INFORMATION

POWER METERS

8651A	Single Input Universal Power Meter
	(includes 1 sensor cable)
8652A	Dual Input Universal Power Meter
	(includes 2 sensor cables)

ACCESSORIES

One manual, one power cord.

POWER METER OPTIONS

- 01 Rack mount kit
- 03 8651A Rear Panel Sensor and Calibrator Connections
- 04 8652A Rear Panel Sensor and Calibrator Connections
- 05 Soft Carry Case
- 07 Side Mounted Carrying Handle
- 08 Transit Case, (Includes Soft Carry Case)
- 09 Dual Rack Mount Kit (with assembly instructions)
- 10 Dual Rack Mount Kit (factory assembled)
- 12 1 GHz, 50 MHz Switchable Calibrator
- 13 8651A Rear Panel Input Connector
- 14 8652A Rear Panel Input Connectors

¹⁰ Depending on sensor used. ¹¹ MAP (Modulated Average Power), PAP (Pulse Average Power), BAP (Burst Average Power). ¹² Specified performance applies with maximum averaging and 24 hour warm-up at constant temperature. ¹³ Operates in Normal Mode only. ¹⁴ Display contrast reduces above 50° C. ¹⁵ Does not apply to 80701A Sensor below 500 MHz.

Specifications subject to change without notice

Giga-tronics

Giga-tronics Incorporated 4650 Norris Canyon Road San Ramon, California 94583 Telephone: 800 726 4442 or 925 328 4650 Telefax: 925 328 4700 Web Site: www.gigatronics.com

